



# California Medical Device Recall Information



## Recall Name

### Volcano Corporation Recalls Pioneer Plus Re-Entry Catheter Due to a Potential Defect

Recall Date	Product Description	Recalling Firm	Recall Reason
9/2/15	Volcano Pioneer Plus Re-Entry Catheter  Product Code: <ul style="list-style-type: none"><li>• PPLUS20</li></ul>	<b>Volcano Corp.</b> Rancho Cordova, CA	<i>Potential that a small wire could extend through the catheter shaft.</i>
Recall Class	Product Identification	Distribution	Affected Dates
II	<b>All Lots</b>	<b>CA</b> , nationwide	Distributed between:  October 7, 2013 and July 16, 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRES/res.cfm?id=139266>